

计量授权证书附件

序号 No.	开展商品量/商品 包装计量检验的参数 名称 Parameters of Metrological Testing for Quantity/Package of Products	测量范围 Measurement Range	准确度等级或 测量扩展不确定度 Accuracy or Expanded Measurement Uncertainty	依据检定 规程编号 Number of Verification Specification Referred to
1	质量标注定量包装商 品净含量 (Q_n)	(0~50) g (50~100) g (100~200) g (200~300) g (300~500) g (500~1000) g (1000~10000) g (10000~15000) g (15000~50000) g	$\leq Q_n \times 1.8\%$ $\leq 0.9\text{ g}$ $\leq Q_n \times 0.9\%$ $\leq 1.8\text{ g}$ $\leq Q_n \times 0.6\%$ $\leq 3\text{ g}$ $\leq Q_n \times 0.3\%$ $\leq 30\text{ g}$ $\leq Q_n \times 0.2\%$	JJF 1070
2	体积标注定量包装商 品净含量 (Q_n)	(0~50) mL (50~100) mL (100~200) mL (200~300) mL (300~500) mL (500~1000) mL (1000~10000) mL (10000~15000) mL (15000~50000) mL	$\leq Q_n \times 1.8\%$ $\leq 0.9\text{ mL}$ $\leq Q_n \times 0.9\%$ $\leq 1.8\text{ mL}$ $\leq Q_n \times 0.6\%$ $\leq 3\text{ mL}$ $\leq Q_n \times 0.3\%$ $\leq 30\text{ mL}$ $\leq Q_n \times 0.2\%$	JJF 1070
3	长度标注定量包装商 品净含量 (Q_n)	$Q_n \leq 5\text{ m}$ $Q_n > 5\text{ m}$	--- $\leq Q_n \times 0.4\%$	JJF 1070
4	面积标注定量包装商 品净含量 (Q_n)	全部 Q_n	$\leq Q_n \times 0.6\%$	JJF 1070
5	计数标注定量包装商 品净含量 (Q_n)	$Q_n \leq 50$ $Q_n > 50$	--- $\leq Q_n \times 0.2\%$	JJF 1070

-----以下空白-----